

DATA SHEET

Water Penetration - Dr Carstens Method



58-E0023. Water penetration test set (Dr Carstens method)

General description

For the determination of water permeability of building materials and prefabricated building parts. The test can be performed either horizontally or vertically. The test set consists of 3 vertical and 3 horizontal, water penetration test tubes, 250 ml washing bottle and 250 g of plasticine cement.

Case dimension: 420x280x90 mm

Weight approx.: 1.75 kg

Standards

- EN 1323

Products

58-E0023

Water penetration test set. Dr. Karstens method.

Accessories and consumables

58-E0023/3

Plasticine cement, 250 g.